

Visy Pulp and Paper Tumut CEMS - Correction Log

Reporting Period: 1/05/2026 - 1/06/2026

Environment Protection Licence No: 10232

Main Stack 1

Monitoring Location No: 1
 Monitoring Type: Continuous
 Sample Type: Air
 Description: Exit point from Stack 1 to atmosphere

Opacity						Period: 6 Minutes	Limit: 20.00 %
Start Time	End Time	Cause	Operational State	Explanation	Max Reading		
01/05/26 09:00	01/05/26 09:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.44		
02/05/26 09:00	02/05/26 09:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.99		
03/05/26 09:00	03/05/26 09:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.82		
04/05/26 09:00	04/05/26 09:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.62		
05/05/26 09:00	05/05/26 09:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.61		
12/05/26 08:54	12/05/26 09:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.48		
13/05/26 08:54	13/05/26 09:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.52		
14/05/26 08:54	14/05/26 09:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.47		
15/05/26 08:54	15/05/26 09:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.78		
16/05/26 08:54	16/05/26 09:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.80		
17/05/26 08:54	17/05/26 09:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.99		
18/05/26 08:54	18/05/26 09:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.24		
19/05/26 08:54	19/05/26 09:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.56		
20/05/26 08:54	20/05/26 09:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.31		
21/05/26 08:54	21/05/26 09:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.75		
22/05/26 08:54	22/05/26 09:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.08		
23/05/26 08:54	23/05/26 09:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.82		
24/05/26 08:54	24/05/26 09:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.25		

25/05/26 08:54	25/05/26 09:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.99
26/05/26 08:54	26/05/26 09:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.23
27/05/26 08:54	27/05/26 09:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.41
28/05/26 08:54	28/05/26 09:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.19
29/05/26 08:54	29/05/26 09:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.51

Power Boiler

Monitoring Location No: 3
Monitoring Type: Continuous
Sample Type: Air
Description: Discharge duct downstream of Power Boiler prior to junction with Stack 1

Nitrogen Oxides (as NO ₂)		Period: 60 Minutes	Limit: 300.00 mg/Nm ³		
Start Time	End Time	Cause	Operational State	Explanation	Max Reading
18/05/26 10:00	18/05/26 12:00	Instrument Error	Normal (Steady State)	Motherboard hardware was playing up - causing false readings	399.08

Authorised By:

Uday Bhagwat
 Pulp Mill Manager



Johan Stoltz
 General Manager

