

Visy Pulp and Paper Tumut CEMS - Correction Log

13/03/2025

Reporting Period:

1/02/2025 - 1/03/2025

Environment Protection Licence No: 10232

Main Stack 1

Monitoring Location No: 1

Monitoring Type Continuous

Sample Type: Air

Description: Exit point from Stack 1 to atmosphere

Opacity		Period: 6 Minutes	Limit: 20.00 %		
Start Time	End Time	Cause	Operational State	Explanation	Max Reading
03/02/25 11:24	03/02/25 11:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.45
04/02/25 11:24	04/02/25 11:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.02
05/02/25 11:24	05/02/25 11:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.67
06/02/25 11:24	06/02/25 11:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.29
07/02/25 11:24	07/02/25 11:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.89
08/02/25 11:24	08/02/25 11:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.18
09/02/25 11:24	09/02/25 11:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.69
10/02/25 11:24	10/02/25 11:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.36
11/02/25 11:24	11/02/25 11:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.11
12/02/25 11:24	12/02/25 11:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.49
13/02/25 11:24	13/02/25 11:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.61
14/02/25 11:24	14/02/25 11:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	24.42
15/02/25 11:24	15/02/25 11:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	24.97
16/02/25 11:24	16/02/25 11:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	24.42
17/02/25 11:24	17/02/25 11:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	25.34
18/02/25 11:24	18/02/25 11:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	24.79
19/02/25 11:24	19/02/25 11:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	25.91
20/02/25 11:24	20/02/25 11:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	25.20

21/02/25 11:24	21/02/25 11:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.31
22/02/25 11:24	22/02/25 11:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.57
25/02/25 11:24	25/02/25 11:30	Lime Kiln A	Equipment Issue/Failure	While calibrating the Lime Density Tx, a faulty value input caused the Kiln to trip.	28.98
27/02/25 11:18	27/02/25 11:24	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.12
28/02/25 11:18	28/02/25 11:24	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.77

Main Stack 2

Monitoring Location No: 22
Monitoring Type: Continuous
Sample Type: Air
Description: Exit point from Stack 2 to atmosphere

Hydrogen Chloride (HCl)		Period: 60 Minutes	Limit: 50.00 mg/Nm3		
Start Time	End Time	Cause	Operational State	Explanation	Max Reading
19/02/25 12:00	19/02/25 13:00	Instrument Error	Equipment Issue/Failure	Lier Siegler working on CEMS system	63.60
19/02/25 15:00	19/02/25 16:00	Instrument Error	Equipment Issue/Failure	Lier Siegler working on CEMS system	53.68

Nitrogen Oxides (as NO2)		Period: 60 Minutes	Limit: 400.00 mg/Nm3		
Start Time	End Time	Cause	Operational State	Explanation	Max Reading
19/02/25 11:00	19/02/25 12:00	Instrument Error	Equipment Issue/Failure	Lier Siegler working on CEMS system	670.15

TRS (as H2S)		Period: 60 Minutes	Limit: 3.60 mg/Nm3		
Start Time	End Time	Cause	Operational State	Explanation	Max Reading
19/02/25 11:00	19/02/25 12:00	Instrument Error	Equipment Issue/Failure	Lier Siegler working on CEMS system	5.06

Authorised By:

Uday Bhagwat
Pulp Mill Manager



Johan Stoltz
General Manager



