

Visy Pulp and Paper Tumut CEMS - Correction Log

12/05/2025

Reporting Period:

1/04/2025 - 1/05/2025

Environment Protection Licence No: 10232

Main Stack 1

Monitoring Location No: 1

Monitoring Type: Continuous

Sample Type: Air

Description: Exit point from Stack 1 to atmosphere

Opacity					
Period: 6 Minutes		Limit: 20.00 %			
Start Time	End Time	Cause	Operational State	Explanation	Max Reading
01/04/25 11:12	01/04/25 11:18	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.40
02/04/25 11:12	02/04/25 11:18	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.85
03/04/25 11:12	03/04/25 11:18	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	24.19
04/04/25 11:12	04/04/25 11:18	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	24.97
05/04/25 11:12	05/04/25 11:18	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	24.93
06/04/25 10:12	06/04/25 10:18	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.98
07/04/25 10:12	07/04/25 10:18	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	25.75
08/04/25 10:12	08/04/25 10:18	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	25.10
09/04/25 10:12	09/04/25 10:18	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	25.33
10/04/25 10:12	10/04/25 10:18	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	24.85
11/04/25 10:12	11/04/25 10:18	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.50
12/04/25 10:12	12/04/25 10:18	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.74
17/04/25 10:06	17/04/25 10:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.27
18/04/25 10:06	18/04/25 10:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.35
19/04/25 10:06	19/04/25 10:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.10
20/04/25 10:06	20/04/25 10:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.42
21/04/25 10:06	21/04/25 10:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.90
22/04/25 10:06	22/04/25 10:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.29

23/04/25 10:06	23/04/25 10:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.69
24/04/25 10:06	24/04/25 10:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.43
25/04/25 10:06	25/04/25 10:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.75
26/04/25 10:06	26/04/25 10:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.48
27/04/25 10:06	27/04/25 10:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.75
28/04/25 10:06	28/04/25 10:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	24.06
29/04/25 10:06	29/04/25 10:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	24.20
30/04/25 10:06	30/04/25 10:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	25.12


Power Boiler

Monitoring Location No: 3
Monitoring Type Continuous
Sample Type: Air
Description: Discharge duct downstream of Power Boiler prior to junction with Stack 1

Carbon Monoxide (CO)		Period: 60 Minutes	Limit: 140.00 mg/Nm ³		
Start Time	End Time	Cause	Operational State	Explanation	Max Reading
07/04/25 08:00	07/04/25 09:00	Instrument Error	Power Boiler Scheduled Start-up/Shut-down	Power boiler under shutdown	195.02

Authorised By:

Uday Bhagwat
Pulp Mill Manager



Johan Stoltz
General Manager

