

Visy Pulp and Paper Tumut CEMS - Correction Log

16/06/2025

Reporting Period: 1/05/2025 - 1/06/2025 Environment Protection Licence No: 10232

Main Stack 1

Monitoring Location No: 1
 Monitoring Type: Continuous
 Sample Type: Air
 Description: Exit point from Stack 1 to atmosphere

Flow		Period: 60 Minutes	Limit: 100.00 Nm3/s		
Start Time	End Time	Cause	Operational State	Explanation	Max Reading
21/05/25 09:00	21/05/25 10:00	Instrument Error	Normal (Steady State)	Preventative maintenance from Group Instrumentation	103.59

Nitrogen Oxides (as NO2)		Period: 60 Minutes	Limit: 400.00 mg/Nm3		
Start Time	End Time	Cause	Operational State	Explanation	Max Reading
21/05/25 09:00	21/05/25 10:00	Instrument Error	Normal (Steady State)	Preventative maintenance from Group Instrumentation	432.82

Opacity		Period: 6 Minutes	Limit: 20.00 %		
Start Time	End Time	Cause	Operational State	Explanation	Max Reading
01/05/25 10:06	01/05/25 10:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	24.40
02/05/25 10:06	02/05/25 10:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	24.81
03/05/25 10:06	03/05/25 10:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	25.97
04/05/25 10:06	04/05/25 10:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	25.34
05/05/25 10:06	05/05/25 10:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	24.72
06/05/25 10:06	06/05/25 10:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.19
12/05/25 10:00	12/05/25 10:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.30
13/05/25 10:00	13/05/25 10:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.95
14/05/25 10:00	14/05/25 10:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.46

15/05/25 10:00	15/05/25 10:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.75
16/05/25 10:00	16/05/25 10:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.91
16/05/25 12:00	16/05/25 12:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.72
17/05/25 10:00	17/05/25 10:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.12
18/05/25 10:00	18/05/25 10:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.67
19/05/25 10:00	19/05/25 10:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.84
20/05/25 10:00	20/05/25 10:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	24.76
21/05/25 10:00	21/05/25 10:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	25.32
22/05/25 10:00	22/05/25 10:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	24.22
23/05/25 10:00	23/05/25 10:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	25.03
24/05/25 10:00	24/05/25 10:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	25.12
25/05/25 10:00	25/05/25 10:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	24.95
26/05/25 10:00	26/05/25 10:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	25.92
27/05/25 10:00	27/05/25 10:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	26.20
28/05/25 10:00	28/05/25 10:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	27.20
29/05/25 10:00	29/05/25 10:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	25.50
30/05/25 10:00	30/05/25 10:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.46
31/05/25 10:00	31/05/25 10:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.48

Authorised By:

Uday Bhagwat
Pulp Mill Manager



Johan Stoltz
General Manager

