

Visy Pulp and Paper Tumut CEMS - Correction Log

Reporting Period: 1/09/2025 - 1/10/2025

Environment Protection Licence No: 10232

10/10/2025

Main Stack 1

Monitoring Location No: 1
 Monitoring Type: Continuous
 Sample Type: Air
 Description: Exit point from Stack 1 to atmosphere

Opacity		Period: 6 Minutes	Limit: 20.00 %		
Start Time	End Time	Cause	Operational State	Explanation	Max Reading
04/09/25 13:06	04/09/25 13:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.50
05/09/25 13:06	05/09/25 13:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.93
06/09/25 13:06	06/09/25 13:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.95
07/09/25 13:06	07/09/25 13:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.07
08/09/25 13:06	08/09/25 13:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.84
09/09/25 13:06	09/09/25 13:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.59
10/09/25 13:06	10/09/25 13:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.31
11/09/25 13:06	11/09/25 13:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.44
12/09/25 13:06	12/09/25 13:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.27
13/09/25 13:06	13/09/25 13:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.11
14/09/25 13:06	14/09/25 13:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.74
15/09/25 13:06	15/09/25 13:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.98
16/09/25 13:06	16/09/25 13:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.53
17/09/25 13:06	17/09/25 13:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.34
18/09/25 13:06	18/09/25 13:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.39
19/09/25 13:06	19/09/25 13:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	24.65
20/09/25 13:06	20/09/25 13:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.46
21/09/25 13:06	21/09/25 13:12	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.30

28/09/25 13:00	28/09/25 13:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.64
29/09/25 13:00	29/09/25 13:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.71
30/09/25 13:00	30/09/25 13:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.82

Authorised By:

Uday Bhagwat
Pulp Mill Manager



Johan Stoltz
General Manager

