



Visy Pulp and Paper Tumut CEMS - Correction Log

2/07/2026

Reporting Period:

1/06/2026 - 1/07/2026

Environment Protection Licence No: 10232

Main Stack 1


Monitoring Location No: 1
 Monitoring Type: Continuous
 Sample Type: Air
 Description: Exit point from Stack 1 to atmosphere

Opacity		Period: 6 Minutes	Limit: 20.00 %		
Start Time	End Time	Cause	Operational State	Explanation	Max Reading
05/06/26 08:48	05/06/26 08:54	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.35
06/06/26 08:48	06/06/26 08:54	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.40
07/06/26 08:48	07/06/26 08:54	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.53
08/06/26 08:48	08/06/26 08:54	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.86
09/06/26 08:48	09/06/26 08:54	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.99
10/06/26 08:48	10/06/26 08:54	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.72
11/06/26 08:48	11/06/26 08:54	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.54
12/06/26 08:48	12/06/26 08:54	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.41
12/06/26 11:12	12/06/26 11:18	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	27.64
13/06/26 08:48	13/06/26 08:54	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.52
14/06/26 08:48	14/06/26 08:54	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.01
15/06/26 08:48	15/06/26 08:54	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.20
16/06/26 08:48	16/06/26 08:54	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.39
17/06/26 08:48	17/06/26 08:54	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.30
18/06/26 08:48	18/06/26 08:54	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.96
19/06/26 08:48	19/06/26 08:54	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.75
20/06/26 08:48	20/06/26 08:54	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.97
21/06/26 08:48	21/06/26 08:54	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.51

22/06/26 08:48	22/06/26 08:54	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.43
29/06/26 08:42	29/06/26 08:48	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.20
30/06/26 08:42	30/06/26 08:48	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.53

Authorised By:

Uday Bhagwat
Pulp Mill Manager



(S. Olivier)

Johan Stoltz
General Manager

