

Visy Pulp and Paper Tumut CEMS - Correction Log

2/09/2026

Reporting Period: 1/08/2026 - 1/09/2026

Environment Protection Licence No: 10232

Main Stack 1

Monitoring Location No: 1

Monitoring Type: Continuous

Sample Type: Air


Description: Exit point from Stack 1 to atmosphere

Opacity Period: 6 Minutes Limit: 20.00 %					
Start Time	End Time	Cause	Operational State	Explanation	Max Reading
01/08/26 08:36	01/08/26 08:42	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.51
02/08/26 08:36	02/08/26 08:42	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.84
03/08/26 08:36	03/08/26 08:42	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.57
04/08/26 08:36	04/08/26 08:42	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.80
05/08/26 08:36	05/08/26 08:42	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.93
06/08/26 08:36	06/08/26 08:42	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.31
07/08/26 08:36	07/08/26 08:42	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.82
07/08/26 09:30	07/08/26 09:36	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	36.52
08/08/26 08:36	08/08/26 08:42	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.47
15/08/26 08:30	15/08/26 08:36	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.14
16/08/26 08:30	16/08/26 08:36	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.33
17/08/26 08:30	17/08/26 08:36	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.66
18/08/26 08:30	18/08/26 08:36	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.68
19/08/26 08:30	19/08/26 08:36	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.72
20/08/26 08:30	20/08/26 08:36	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.87
21/08/26 08:30	21/08/26 08:36	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.28
22/08/26 08:30	22/08/26 08:36	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.08
23/08/26 08:30	23/08/26 08:36	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.90

24/08/26 08:30	24/08/26 08:36	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.94
25/08/26 08:30	25/08/26 08:36	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.22
26/08/26 08:30	26/08/26 08:36	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.38
27/08/26 08:30	27/08/26 08:36	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.86
28/08/26 08:30	28/08/26 08:36	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.15
29/08/26 08:30	29/08/26 08:36	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.56
30/08/26 08:30	30/08/26 08:36	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.38
31/08/26 08:30	31/08/26 08:36	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.98

Authorised By:

Uday Bhagwat
Pulp Mill Manager



Johan Stoltz
General Manager


 PP RAMN DAYAN AND
 04/09/26