VISY
FOR A BETTER WORLD

Visy Pulp and Paper Tumut CEMS - Correction Log

7/10/2026

Reporting Period: 1/09/2026 - 1/10/2026 **Environment Protection Licence No:** 10232

Main Stack 1

Monitoring Location No: 1
Monitoring Type Continuous
Sample Type: Air
Description: Exit point from Stack 1 to atmosphere

Hydrogen Chloride (HCl) **Period:** 60 Minutes **Limit:** 50.00 mg/Nm3

Start Time	End Time	Cause	Operational State	Explanation	Max Reading
14/09/26 13:00	14/09/26 14:00	Instrument Error	Normal (Steady State)	PROCAL went into fault mode	134.11

Opacity **Period:** 6 Minutes **Limit:** 20.00 %

Start Time	End Time	Cause	Operational State	Explanation	Max Reading
01/09/26 08:30	01/09/26 08:36	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.21
03/09/26 09:24	03/09/26 09:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	28.87
08/09/26 08:24	08/09/26 08:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.23
09/09/26 08:24	09/09/26 08:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.80
10/09/26 08:24	10/09/26 08:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.15
11/09/26 08:24	11/09/26 08:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.47
12/09/26 08:24	12/09/26 08:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.79
13/09/26 08:24	13/09/26 08:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.38
14/09/26 08:24	14/09/26 08:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.62
15/09/26 08:24	15/09/26 08:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.89
16/09/26 08:24	16/09/26 08:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.44
17/09/26 08:24	17/09/26 08:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.83
18/09/26 08:24	18/09/26 08:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.65

19/09/26 08:24	19/09/26 08:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.97
20/09/26 08:24	20/09/26 08:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.34
21/09/26 08:24	21/09/26 08:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.80
22/09/26 08:24	22/09/26 08:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.49
23/09/26 08:24	23/09/26 08:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.30
24/09/26 08:24	24/09/26 08:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.46
25/09/26 08:24	25/09/26 08:30	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.03

Authorised By:

Uday Bhagwat
Pulp Mill Manager

Johan Stoltz
General Manager