



Visy Pulp and Paper Tumut CEMS - Correction Log

12/05/2022

Reporting Period: 1/04/2022 - 1/05/2022

Environment Protection Licence No: 10232

Main Stack 1

Monitoring Location No: 1
 Monitoring Type: Continuous
 Sample Type: Air
 Description: Exit point from Stack 1 to atmosphere

Opacity		Period: 6 Minutes	Limit: 20.00 %			
Start Time	End Time	Cause	Operational State	Explanation	Max Reading	
01/04/22 18:00	01/04/22 18:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	25.52	
02/04/22 18:00	02/04/22 18:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	25.32	
03/04/22 17:00	03/04/22 17:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	25.57	
04/04/22 17:00	04/04/22 17:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	26.77	
05/04/22 12:48	05/04/22 13:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	34.80	
05/04/22 17:00	05/04/22 17:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.88	
06/04/22 17:00	06/04/22 17:06	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.06	
12/04/22 16:54	12/04/22 17:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.73	
13/04/22 16:54	13/04/22 17:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.65	
14/04/22 16:54	14/04/22 17:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.91	
15/04/22 16:54	15/04/22 17:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.43	
16/04/22 16:54	16/04/22 17:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.79	
17/04/22 16:54	17/04/22 17:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.15	
18/04/22 16:54	18/04/22 17:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.10	
19/04/22 16:54	19/04/22 17:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.22	
20/04/22 16:54	20/04/22 17:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.16	
21/04/22 16:54	21/04/22 17:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.97	
22/04/22 16:54	22/04/22 17:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.41	

23/04/22 16:54	23/04/22 17:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.35
24/04/22 16:54	24/04/22 17:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.03
25/04/22 16:54	25/04/22 17:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	25.42
26/04/22 16:54	26/04/22 17:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.01
27/04/22 16:54	27/04/22 17:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.86
28/04/22 16:54	28/04/22 17:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	24.90
29/04/22 16:54	29/04/22 17:00	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.16

Authorised By:

Uday Bhagwat
Pulp Mill Manager

Johan Stoltz
General Manager

ALTAIN ZOLIO 

