



Visy Pulp and Paper Tumut CEMS - Correction Log

4/07/2022

Reporting Period: 1/06/2022 - 1/07/2022

Environment Protection Licence No: 10232

Main Stack 1

Monitoring Location No: 1
 Monitoring Type: Continuous
 Sample Type: Air
 Description: Exit point from Stack 1 to atmosphere

Opacity						Period: 6 Minutes	Limit: 20.00 %
Start Time	End Time	Cause	Operational State	Explanation	Max Reading		
01/06/22 16:42	01/06/22 16:48	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.72		
02/06/22 16:42	02/06/22 16:48	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.15		
03/06/22 16:42	03/06/22 16:48	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.90		
04/06/22 16:42	04/06/22 16:48	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.03		
05/06/22 16:42	05/06/22 16:48	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.40		
06/06/22 16:42	06/06/22 16:48	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.87		
07/06/22 16:42	07/06/22 16:48	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.20		
08/06/22 16:42	08/06/22 16:48	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.86		
09/06/22 16:42	09/06/22 16:48	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.91		
10/06/22 16:42	10/06/22 16:48	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.74		
11/06/22 16:42	11/06/22 16:48	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.99		
12/06/22 16:42	12/06/22 16:48	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.85		
13/06/22 16:42	13/06/22 16:48	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.95		
14/06/22 16:42	14/06/22 16:48	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	25.62		
15/06/22 16:42	15/06/22 16:48	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	25.37		
16/06/22 16:42	16/06/22 16:48	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.13		
17/06/22 16:42	17/06/22 16:48	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.64		
23/06/22 16:36	23/06/22 16:42	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.55		

24/06/22 16:36	24/06/22 16:42	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.04
25/06/22 16:36	25/06/22 16:42	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.98
26/06/22 16:36	26/06/22 16:42	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.94
27/06/22 16:36	27/06/22 16:42	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.51
28/06/22 16:36	28/06/22 16:42	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.15
29/06/22 16:36	29/06/22 16:42	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.29
30/06/22 16:36	30/06/22 16:42	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.08

Authorised By:

Uday Bhagwat
Pulp Mill Manager



Johan Stoltz
General Manager

