

Visy Pulp and Paper Tumut CEMS - Correction Log

6/10/2021

Reporting Period:

1/09/2021 - 1/10/2021

Environment Protection Licence No:

10232

Main Stack 1

Monitoring Location No: 1
 Monitoring Type: Continuous
 Sample Type: Air
 Description: Exit point from Stack 1 to atmosphere

Opacity					
Period: 6 Minutes		Limit: 20.00 %			
Start Time	End Time	Cause	Operational State	Explanation	Max Reading
01/09/21 12:48	01/09/21 12:54	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.26
02/09/21 12:48	02/09/21 12:54	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.38
03/09/21 12:48	03/09/21 12:54	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.04
04/09/21 12:48	04/09/21 12:54	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.30
05/09/21 12:48	05/09/21 12:54	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.65
06/09/21 12:48	06/09/21 12:54	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.94
07/09/21 12:48	07/09/21 12:54	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.50
08/09/21 12:48	08/09/21 12:54	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.32
09/09/21 12:06	09/09/21 12:18	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	31.19
09/09/21 12:48	09/09/21 12:54	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.58
10/09/21 12:48	10/09/21 12:54	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.18
11/09/21 12:48	11/09/21 12:54	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	25.01
12/09/21 12:48	12/09/21 12:54	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	25.06
13/09/21 12:48	13/09/21 12:54	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	25.41
14/09/21 12:48	14/09/21 12:54	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	24.72
15/09/21 12:48	15/09/21 12:54	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	25.15
16/09/21 12:48	16/09/21 12:54	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.01
17/09/21 12:48	17/09/21 12:54	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	23.50


18/09/21 12:48	18/09/21 12:54	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.45
23/09/21 12:42	23/09/21 12:48	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.77
24/09/21 12:42	24/09/21 12:48	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	20.78
25/09/21 12:42	25/09/21 12:48	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.26
26/09/21 12:42	26/09/21 12:48	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.74
27/09/21 12:42	27/09/21 12:48	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.13
28/09/21 12:42	28/09/21 12:48	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	21.69
29/09/21 12:42	29/09/21 12:48	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.45
30/09/21 12:42	30/09/21 12:48	Auto Zero Span Verification	Normal (Steady State)	Auto Calibration	22.81

Main Stack 2

Monitoring Location No: 22
Monitoring Type: Continuous
Sample Type: Air
Description: Exit point from Stack 2 to atmosphere

Opacity						Period: 6 Minutes	Limit: 20.00 %
Start Time	End Time	Cause	Operational State	Explanation	Max Reading		
07/09/21 09:36	07/09/21 09:42	Instrument Error	Normal (Steady State)	CEMS Calibrations done by E&I	23.55		

Authorised By:

Uday Bhagwat
Pulp Mill Manager


Johan Stoltz
General Manager
